

Können die Alarmsignale verändert werden?

In the SENSORS menu section of the SmartXcan web interface, the alarm settings can be adjusted. Here you can select the type of signal (continuous, pulsing) and the signal duration (1 sec. to 1 hour).

Alarm settings		
Signal type	Pulsating	Continuous
Signal duration	1 sec to 1 hour	1 sec to 1 hour